

INSPECTION CERTIFICATE 3.1 (Cont.)

**Inspection Report for Test Sieve
in accordance with DIN ISO 3310-1**

Apertures Warp				Apertures Weft				Wire Diameter			
1,003	0,991			1,006	1,007			Warp			
1,007	1,016			1,005	1,003			0,541	0,547	0,542	
1,023	0,996			1,008	0,998			0,546	0,545	0,545	
1,023	1,032			1,005	1,002			0,540	0,540	0,546	
1,012	1,003			1,000	1,004			0,540	0,538	0,582	
1,012	1,023			1,002	1,006			0,540	0,546	0,548	
0,951	0,991			0,993	0,991			0,558	0,540	0,546	
1,033	1,009			1,009	0,996			0,545	0,551	0,544	
1,037	1,016			1,006	1,009			0,545	0,548	0,550	
0,995	1,031			1,004	0,959			0,573	0,549	0,538	
0,996	0,978			1,016	0,990			0,548	0,534	0,547	
1,035	0,998			1,000	0,992			0,542	0,541	0,539	
1,009	1,002			1,012	1,012			0,547	0,543	0,544	
1,017	1,020			0,987	1,000			0,542	0,540	0,545	
1,016	1,007			1,009	1,001			0,545	0,545	0,542	
1,001	1,034			1,005	0,993			0,546	0,544	0,542	
1,018	0,993			0,993	1,005			0,547	0,549	0,544	
1,003	1,025			1,009	1,014			0,543	0,542	0,544	
1,018	1,007			1,005	1,012			0,550	0,541	0,540	
1,012	1,028			1,002	0,996			0,545	0,546	0,546	
1,002	0,999			0,994	1,000			0,547	0,543		
1,018	1,006			1,010	1,007			0,540	0,547		
1,010	1,002			0,985	1,003			Weft			
1,017	1,014			1,007	0,990			0,542	0,543	0,547	
1,010	0,990			1,001	1,006			0,545	0,549	0,544	
1,010	1,017			0,996	0,990			0,546	0,546	0,544	
1,013	1,000			1,009	1,014			0,551	0,545	0,547	
1,018	1,016			1,004	1,010			0,546	0,546	0,547	
1,000	0,997			0,993	0,990			0,547	0,542	0,549	
1,025	1,026			1,002	1,004			0,544	0,546	0,546	
0,986	0,998			1,011	1,006			0,545	0,545	0,548	
1,033	1,027			0,995	1,008			0,547	0,545	0,548	
1,017	1,012			0,995	1,000			0,546	0,544	0,545	
1,008	1,004			1,012	0,994			0,547	0,552	0,551	
1,002	0,985			1,003	1,019			0,543	0,542	0,545	
1,031	1,023			0,997	1,002			0,556	0,544	0,548	
0,997	1,003			1,005	0,998			0,543	0,550	0,547	
1,015	1,021			0,988	1,007			0,546	0,543	0,544	
1,000	0,990			1,001	1,003			0,547	0,555	0,548	
1,034	1,012			1,010	0,997			0,548	0,545	0,590	
1,005	0,988			1,007	1,008			0,547	0,543	0,541	
1,027	1,025			0,991	1,006			0,545	0,549	0,552	
1,000				1,007	1,015			0,546	0,546		
1,010				1,010				0,543	0,546		
0,996				0,997				Mean Values			
1,017				1,001				Warp			
0,988				0,999				0,5453			
1,028				1,001				Weft			
0,993				0,998				0,5470			
0,977				1,005				Sieve Serial Number			
Inspected:	92			Inspected:	93			000968			
Mean \bar{w}	1,0094			Mean \bar{w}	1,0019						
Max w :	1,037			Max w :	1,019						